



test parameters	column: CHIROBIOTIC V2, 10 cm x 2.1 mm I.D., 5 µm particles (15018AST)
	mobile phase: 13 mM ammonium acetate in methanol
	flow rate: 0.2 mL/min
	column temp.: 25 °C
	detector: UV, 230 nm
	injection: 5 µL
	sample: 1 mg/mL in methanol