



test parameters	column: CHIROBIOTIC V2, 25 cm x 4.6 mm I.D., 5 µm particles (15024AST)
	mobile phase: 15 mM ammonium acetate in methanol
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: UV, 215 nm
	injection: 5 µL
	sample: 1 mg/mL in methanol