



test parameters	column: CHIROBIOTIC T2, 25 cm x 4.6 mm I.D., 5 µm particles (16024AST)
	mobile phase: (A) 10 mM ammonium acetate, pH 5.5; (B) acetonitrile; (70:30, A:B)
	flow rate: 0.9 mL/min
	column temp.: 25 °C
	detector: UV, 230 nm
	injection: 2 µL
	sample: 1 mg/mL in methanol