



test parameters	column: CYCLOBOND I 2000, 25 cm x 4.6 mm I.D., 5 µm particles (20024AST)
	mobile phase: (A) acetonitrile; (B) methanol; (C) acetic acid; (D) triethylamine; (90:10:0.3:0.2, A:B:C:D)
	flow rate: 1 mL/min
	column temp.: ambient
	detector: UV, 230 nm
	injection: 1 µL
	sample: 5 mg/mL in methanol