



test parameters	<p>column: CYCLOBOND I 2000 AC, 25 cm x 4.6 mm I.D., 5 µm particles (20124AST)</p> <p>mobile phase: (A) 1.0% triethylamine acetate, pH 4.1; (B) acetonitrile; (85:15, A:B)</p> <p>flow rate: 1 mL/min</p>
	column temp.: 22 °C
	detector: UV, 254 nm
	injection: 2 µL
	<p>sample: 1 mg/mL in methanol</p> <p>peak 1: 9.48 min.</p> <p>peak 2: 12.48 min.</p>