



test parameters	column: CYCLOBOND I 2000 RSP, 25 cm x 4.6 mm I.D., 5 µm particles (20324AST)
	mobile phase: (A) 1% triethylamine acetate, pH 4.1; (B) acetonitrile; (80:20, A:B)
	flow rate: 2 mL/min
	column temp.: ambient
	detector: UV, 230 nm
	injection: 5 µL
	sample: 5 mg/mL in methanol