



test parameters	column: CYCLOBOND I 2000 DMP, 25 cm x 4.6 mm I.D., 5 µm particles (20724AST)
	mobile phase: (A) 0.1% triethylamine acetate, pH 5.0; (B) acetonitrile; (80:20, A:B)
	flow rate: 1.2 mL/min
	column temp.: 23 °C
	detector: UV, 254 nm
	injection: 5 µL
	sample: 5 mg/mL in methanol