



test parameters	column: CHIROBIOTIC T2 guard column, 2 cm x 4 mm I.D., 5 µm particles (16100AST (requires holder 21150AST))
	column: CHIROBIOTIC V2, 15 cm x 4.6 mm I.D., 5 µm particles (15023AST)
	mobile phase: (A) 20 mM ammonium acetate, pH 4.1; (B) methanol; (7:93, A:B)
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: UV, 230 nm
	injection: 2 µL

