



test parameters	column: (R,R) P-CAP-DP, 25 cm × 4.6 mm I.D., 5 µm particles (25024AST)
	mobile phase: (A) hexane; (B) ethanol; (80:20, A:B)
	flow rate: 1.0 mL/min
	column temp.: 23 °C
	detector: UV, 235 nm
	injection: 0.2 mL
	sample: 4 µg/mL