



test parameters	column: Discovery C18, 15 cm × 4.6 mm I.D., 5 µm particles (504955)
	mobile phase: 80:20 water 10 mM ammonium acetate, pH 3.0 with formic acid):acetonitrile
	flow rate: 1.0 mL/min
	column temp.: 35 °C
	detector: UV, 254 nm
	injection: 10 µL
	sample: Cylert (50 µg/mL) and Ritalin (100 µg/mL) in 85:15 water:CH <sub>3</sub> OH