



test parameters	column: Ascentis Express Phenyl-Hexyl, 15 cm x 2.1 mm, 2.7 μ m particles (53338-U)
	mobile phase: (A) 70:30, water with 0.1% TFA:acetonitrile with 0.09% TFA; (B) 55:45, water with 0.1% TFA:acetonitrile with 0.09% TFA
	gradient: 0 to 100% B in 7.5 min
	flow rate: 0.3 mL/min
	pressure: 3960 psi
	column temp.: 30 $^{\circ}$ C
	detector: UV, 215 nm