



test parameters	column: Ascentis Express C18, 10 cm x 4.6 mm, 2.7 µm particles (53827-U)
	mobile phase: (A) water; (B) acetonitrile; (40:60 v/v, A:B)
	flow rate: 2 mL/min
	pressure: 323 bar
	detector: UV, 360 nm
	injection: 10 µL
	sample: carb Method 1004 DNPH Mix 2 (47651-U) dilute 3:2 with water