



test parameters	column: Ascentis Express C18, 15 cm × 4.6mm, 2.7 µm particles (53829-U)
	mobile phase: (A) water, (B) acetonitrile:0.5% triethylamine in water, pH=7.0 (80:20)
	flow rate: 0.8 mL/min
	column temp.: ambient
	detector: UV, 285 nm
	injection: 15 µL