



test parameters	column: Ascentis Express C8, 5 cm x 3 mm I.D., 2.7 µm particles (53848-U)
	mobile phase: (A) 0.1% formic acid in water; (B) 0.1% formic acid in acetonitrile
	gradient: 15 to 100% B in 8.5 min
	flow rate: 0.6 mL/min
	column temp.: 35 °C
	detector: UV, 214 nm
	injection: 2 µL