



test parameters	column: Ascentis Express HILIC, 5 cm x 2.1 mm I.D., 2.7 µm particles (53934-U)
	mobile phase: 5 mM ammonium formate in 95:5 (v/v) acetonitrile:water
	flow rate: 0.2 mL/min
	pressure: 22.6 bar
	column temp.: 35 °C
	detector: MS, ESI(-), SIR mode
	injection: 2 µL