



test parameters	column: Ascentis Express HILIC, 15 cm x 4.6 mm I.D., 2.7 μ m particles (53981-U)
	mobile phase: (A) 13 mM ammonium acetate in 10:90 (v/v) water: acetonitrile; (B) 13 mM ammonium acetate in water
	gradient: held at 0% B for 1 min; 0 to 90% B in 19 min
	flow rate: 1.0 mL/min
	column temp.: 35 $^{\circ}$ C
	detector: ESI(+)
	injection: 10 μ L