



test parameters	column: apHera NH2, 15 cm × 4.6 mm I.D., 5 μm particles (56401AST)
	mobile phase: (A) water: (B) acetonitrile; (20:80, A:B)
	flow rate: 1.0 mL/min
	column temp.: 25 °C
	detector: ELSD, 45 °C, 3.5 psi nitrogen
	injection: 10 μL
	sample: 500 μg/mL in 30:70, water: acetonitrile