



test parameters	column: Discovery BIO Wide Pore C18, 15cm × 10mm I.D., 10µm particles (567208-U)
	mobile phase: (A) (80:20), 0.1% TFA in water:0.1% TFA in acetonitrile; (B) (66:34), 0.1% TFA in water:0.1% TFA in acetonitrile
	flow rate: 4.7 mL/min
	column temp.: 30 °C
	detector: UV, 215 nm
	injection: 24.5 µL
	sample: Cat. No. P2693