



test parameters	column: Discovery BIO Wide Pore C18, 15 cm × 4.6 mm I.D., 5 μm particles (568222-U)
	mobile phase: (A) (10 mM NH ₄ OAc, pH 7):(50% CH ₃ CN in 20 mM NH ₄ OAc, pH 7), 65:35 ; (B) (10 mM NH ₄ OAc, pH 7):(50% CH ₃ CN in 20 mM NH ₄ OAc, pH 7), 25:75
	gradient: 0-100% B in 12.5 min
	flow rate: 1 mL/min
	column temp.: ambient
	detector: UV, 254 nm
	injection: 6 μL(10 μg each) in water