



test parameters	column: Ascentis ES Cyano, 15 cm x 4.6 mm I.D., 5 µm particles (577306-U)
	mobile phase: (A) 10 mM potassium phosphate dibasic, pH 7 with phosphoric acid; (B) acetonitrile; (C) methanol; (25:60:15, A:B:C)
	flow rate: 2 mL/min
	column temp.: 25 °C
	detector: UV, 215 nm
	injection: 5 µL
	sample: 100 µg/mL in water: methanol (70:30)