



test parameters	column: Ascentis C18, 15 cm x 2.1 mm I.D., 5 µm particles (581304-U)
	mobile phase: (A) 25 mM formic acid in water; (B): 50:50, 25 mM formic acid in water:25 mM formic acid in acetonitrile
	flow rate: 0.2 mL/min
	column temp.: 35 °C
	detector: ESI (+), full scan injection: 10 µL
	sample: tryptic digest of carboxymethylated protein mixture 2.4 g/L in 0.1% formic acid