



test parameters	column: Ascentis C18, 10 cm x 4.6 mm I.D., 3 μm particles (581321-U)
	mobile phase: (A) 10 mM ammonium acetate (pH 4.5 with acetic acid); (B) acetonitrile; (25:75, A:B)
	flow rate: 1 mL/min
	column temp.: 35 °C
	detector: UV, 220 nm
	injection: 10 μL
	sample: 30 μg/mL each in mobile phase