



| | |
|-----------------|--|
| test parameters | column: Ascentis C18, 15 cm x 4.6 mm I.D., 5 µm particles (581324-U) |
| | mobile phase: (A) 20 mM phosphoric acid (pH 2.0, unadjusted); (B) acetonitrile; (85:15, A:B) |
| | flow rate: 1 mL/min |
| | column temp.: 35 °C |
| | detector: UV, 220 nm |
| | injection: 10 µL |
| | sample: 50 µg/mL each in mobile phase |