



test parameters	column: Ascentis C18, 25 cm × 4.6 mm I.D., 5 µm particles (581325-U)
	mobile phase: (A) water:formic acid (9:1), (B) acetonitrile:water:formic acid (5:4:1)
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: VIS, 518 nm
	injection: 10 µL
	sample: red wine diluted with water, centrifuged and passed through a 0.45 µm filter