



| | |
|-----------------|---|
| test parameters | column: Ascentis C18, 10 cm x 2.1 mm I.D., 5 µm particles (581326-U) |
| | mobile phase: (A) 20 mM formic acid in water; (B) (50:50), 20 mM formic acid in water:20 mM formic acid in acetonitrile |
| | flow rate: 0.2 mL/min |
| | column temp.: 22 °C |
| | detector: ESI (+), full scan |
| | injection: 2 µL |