


1. Propyl gallate (10 µg/mL)
2. Ethoxyquin (10 µg/mL)
3. 2,4,5-trihydroxybutyrophenone (10 µg/mL)
4. tert-butylhydroquinone (10 µg/mL)
5. Nordihydroguaiaretic acid (10 µg/mL)
6. 3-tert-butyl-4-hydroxyanisole (10 µg/mL)
7. 2,6-di-tert-butyl-4-hydroxymethylphenol (10 µg/mL)
8. Octyl gallate (10 µg/mL)
9. Lauryl gallate (10 µg/mL)
10. 3,5-di-tert-butyl-4-hydroxytoluene (10 µg/mL)

Gradient Program		
Time (min)	%A	%B
0	70	30
10	0	100
15	0	100



test parameters	column: Ascentis C8, 15 cm x 4.6 mm I.D., 5 µm particles (581424-U)
	mobile phase: (A) 5% acetic acid; (B) (50:50) methanol:acetonitrile
	flow rate: 2.0 mL/min
	column temp.: 30 °C
	detector: UV, 280 nm
	injection: 10 µL
	sample: as indicated in 50:50 acetonitrile:2-propanol