



test parameters	column: Ascentis Si, 15 cm × 2.1 mm I.D., 5 μm particles (581509-U)
	mobile phase: (A) 100 mM ammonium formate (pH 3.0, with formic acid), (B) water, (C) acetonitrile
	flow rate: 0.3 mL/min
	column temp.: 35 °C
	detector: ESI (+), full scan
	injection: 2 μL
	sample: in 10:90, (50 mM ammonium formate/formic acid, pH 3.0):acetonitrile