



test parameters	column: Ascentis Si, 5 cm x 4.6 mm I.D., 5 µm particles (581511-U)
	mobile phase: (A) 0.1% ammonium acetate in water (pH unadjusted): (B) 0.1% ammonium acetate in acetonitrile; (10:90, A:B)
	flow rate: 1 mL/min
	column temp.: 35 °C
	detector: UV, 254 nm
	injection: 10 µL
	sample: as indicated in 0.1% ammonium acetate in 10:90:water:acetonitrile