



test parameters	column: Ascentis Phenyl, 15 cm x 4.6 mm I.D., 3 µm particles (581610-U)
	mobile phase: (A) water: (B) acetonitrile; (75:25, A:B)
	flow rate: 1.0 mL/min
	column temp.: 35 °C
	detector: UV, 240 nm
	injection: 10 µL
	sample: 50 µg/mL in 85:15, water:methanol