



test parameters	column: Ascentis Phenyl, 15 cm x 2.1 mm I.D., 5 µm particles (581613-U)
	mobile phase: 10 mM ammonium formate (pH 3.0 with formic acid)
	flow rate: 0.2 mL/min
	column temp.: 35 °C
	detector: UV, 270 nm; 750 psi back pressure regulator on outlet of flow cell
	injection: 1 µL
	sample: as indicated, in mobile phase