


9-10 Ethoxy Groups
(Octylphenol)



Gradient Program

Time (min)	%B
0	30
8	50
15	60
25	80
35	100

test parameters	column: SUPELCOSIL LC-Diol, 25 cm × 4.6 mm I.D., 5 µm particles (58201)
	mobile phase: (A) hexane:methylene chloride (95:5); (B) hexane:methylene chloride:methanol (50:40:10)
	flow rate: 1 mL/min column temp.: 35 °C detector: UV, 280 nm
	injection: 10 µL methylene chloride containing 20 µg surfactant