



test parameters	column: SUPELCOSIL LC-8, 15 cm × 4.6 mm I.D., 5 µm particles (58220-U)
	mobile phase: (A) 1% acetic acid/CH ₃ OH: (B) 1% acetic acid/water
	gradient: 35:65 to 100:0 (20 min), return to 35:65 (5 min)
	flow rate: 1.5 mL/min
	column temp.: 30 °C
	detector: UV, 280 nm