



test parameters	column: SUPELCOSIL LC-18, 25 cm × 4.6 mm I.D., 5 μm particles (58298)
	mobile phase: (A) acetonitrile; (B) acetone; (59.0:41.0, v/v, A:B)
	flow rate: 1 mL/min
	detector: RI
	injection: 10 μL of 9% C8 to C18 saturated and unsaturated FAMES in mobile phase