



| | |
|-----------------|--|
| test parameters | column: SUPELCOSIL LC-PAH, 15 cm x 4.6 mm I.D., 5 µm particles (58318) |
| | mobile phase: (A) water, (B) methanol |
| | gradient: 55 to 100% B in 5 min; held at 100%B for 10 min |
| | flow rate: 1.5 mL/min |
| | column temp.: 35 °C |
| | detector: UV, 254 nm |
| | injection: 5 µL |