



test parameters	column: SUPELCOSIL LC-NH2, 25 cm × 4.6 mm I.D., 5 µm particles (58338)
	mobile phase: (A) acetonitrile; (B) water; (75:25, A:B)
	flow rate: 2 mL/min
	detector: RI
	injection: 10 µL, 150 µg each sugar