



test parameters	column: SUPELCOSIL LC-PCN, 15cm × 4.6mm I.D., 5µm particles (58377)
	mobile phase: (A) acetonitrile: (B) methanol: (C) 0.01M dipotassium hydrogen phosphate, pH 7.0 with phosphoric acid (60:15:25, A:B:C)
	flow rate: 2 mL/min
	column temp.: 30 °C
	detector: UV, 215 nm
	injection: 50ng each drug