



test parameters	column: SUPELCOSIL LC-18-S, 15 cm × 4.6 mm I.D., 5 μm particles (58931)
	mobile phase: 0.05 M K ₂ HPO ₄ /KH ₂ PO ₄ , pH 4.0:methanol (A) 97.5:2.5; (B) 80:20
	flow rate: 1 mL/min
	column temp.: 30 °C
	detector: UV, 254 nm
	sample: nucleoside standards in water