



test parameters	column: SUPELCOSIL LC-18-T, 15 cm × 4.6 mm I.D., 3 μm particles (58970-U)
	mobile phase: (A) 0.1M potassium phosphate buffer/4mM tertrabutylammonium hydrogen sulfate, pH6.0; (B) A:methanol (70:30), pH 7.2
	flow rate: 1.5 mL/min
	detector: UV, 254 nm
	injection: 20 μL deionized water (25-530 μM each analyte)