



test parameters	column: SUPELCOSIL LC-8, 15 cm × 4.6 mm I.D., 3 µm particles (58983)
	mobile phase: acetonitrile:tetrahydrofuran: 0.1% phosphoric acid (50.4:21.6:28)
	flow rate: 1 mL/min
	column temp.: 35 °C
	detector: UV, 215 nm
	injection: 10 µL