



test parameters	column: SUPELCOGEL Ca, 30 cm × 7.8 mm I.D. (59305-U)
	mobile phase: water
	flow rate: 1.5 mL/min
	column temp.: 80 °C
	detector: RI
	injection: 10 µL, 2.5 mg/mL each analyte in water