



test parameters	column: SUPELCOGEL Ag, 30 cm × 7.8 mm I.D. (59315)
	mobile phase: water
	flow rate: 0.5 mL/min
	column temp.: 85 °C
	detector: RI
	injection: 10 µL of 1g syrup/10 mL DI water, filtered (0.20 µm filter)