



test parameters	column: Discovery BIO PolyMA-SCX, 5cm × 4.6mm, 5µm particles (59601-U)
	mobile phase: (A) 20 mM sodium phosphate, pH 7.0; (B) 20 mM sodium phosphate, 0.5 M NaCl, pH 7.0
	gradient: 0 to 100% B in 30 min. (linear)
	flow rate: 0.5 mL/min
	column temp.: 25 °C
	detector: UV, 280 nm
	injection: 10 µL