



test parameters	column: Discovery BIO Wide Pore C18, 15 cm × 0.5 mm I.D., 5 μm particles (65519-U)
	mobile phase: (A) 0.1% TFA in water; (B) 0.1% TFA in acetonitrile
	gradient: 5-40% B in 70 min
	flow rate: 14 μL/min
	column temp.: 30 °C
	detector: +ESI mode
	injection: 500 pmol (5μL) β-Lactoglobulin tryptic digest in 50 mM ammonium bicarbonate