



test parameters	column: TSKgel G-Oligo-PW, 30 cm × 7.8 mm I.D., 6 μm particles (808031)
	mobile phase: distilled water
	flow rate: 1 mL/min
	column temp.: 60 °C
	detector: RI