



test parameters	<p>column: TSKgel Ether-5PW, 7.5 cm \times 7.5 mm I.D., 10 μm particles (808641)</p> <p>mobile phase: 0.1 M phosphate buffer, pH 7.0 plus ammonium sulfate, 1.8 M to 0 M in 60 min</p> <p>flow rate: 1 mL/min</p>
	detector: UV, 280 nm