



test parameters	column: TSKgel DEAE-NPR, 3.5 cm × 4.6 mm I.D., 2.5 µm particles (with 5 mm DEAE-NPR guard column) (813075)
	mobile phase: (A) 1 M NaCl/25 mM Tris-HCl, pH 9.0; (B) 25 mM Tris-HCl, pH 9.0
	flow rate: 1 mL/min
	column temp.: ambient
	detector: UV, 260 nm
	injection: 20 µL of 50 µg/mL <i>Hae</i> III digest of pBR322 DNA