



test parameters	column: TSKgel Butyl-NPR, 3.5 cm \times 4.6 mm I.D., 2.5 μ m particles (814947)
	mobile phase: (A) 0.1 M phosphate buffer, pH 7.0 plus ammonium sulfate, 2.3 M to 0 M in 12 min (B & C) = 0.1 M phosphate buffer, pH 7.0 plus ammonium sulfate, 1.8 M to 0 M in 60 min
	flow rate: 1 mL/min
	detector: UV, 280 nm